

## Survey for Dengue Virus NS1 Antigen Testing (DENVAg)

EQA Provider: ESfEQA GmbH  
Heidelberg

ESfEQA Head of Operations: Dr. D. Groche

### Instructions for Use

#### Notes:

The controls do not contain human material, but other biological materials. Therefore, the usual precautions in the laboratory for potentially hazardous material apply for these samples. The samples should be handled by trained personnel only.

Various tasks of the proficiency testing scheme may occasionally be given over to qualified subcontractors. However, ESfEQA is responsible to the participants for the subcontractor's work.

Results from sample analysis may only be disclosed to colleagues from other laboratories after the testing period has been concluded.

By registering for and participating in this EQA participants agree to accept the general terms and conditions of ESfEQA GmbH. These can be retrieved online at [www.esfeqa.eu](http://www.esfeqa.eu).

#### 1. Intended Use

The samples are intended for use as qualitative control material for External Quality Assessment (EQA) in medical laboratories for the following parameter:

Dengue Virus NS1 Antigen

#### 2. Product Description

Sample 1: DENVAg\_2024\_01\_a  
Sample 2: DENVAg\_2024\_01\_b

The lyophilized samples simulate human serum specimens. Positive samples contain recombinant NS1 protein.

#### 3. Storage and Stability

The samples should be stored upright at 2-8 °C. They are stable at least until the deadline for data submission as indicated below.

#### 4. Sample Preparation and Analysis

In order to ensure that all lyophilized sample material is at the bottom of the tube, briefly spin the tube with the lyophilized control (e.g. 30 seconds at 14.000 rpm) or tap the lyophilisate to the bottom of the tube before opening. Reconstitute each sample by adding 0.3 mL of H<sub>2</sub>O (preferably ddH<sub>2</sub>O) or diluent included in your test kit. Allow the control to reconstitute for 5 - 10 minutes, gently swirl and invert the closed tube to ensure homogeneity.

After reconstitution of the samples, they should be handled/treated like patient samples, and testing should be conducted in accordance with the instructions of the instrument and reagent manufacturers.

#### 5. Dates and Submission of Test Results

Testing Period for Sample 1 and 2: 23/04/24 - 14/05/24

For the evaluation of survey results, please indicate the reagent and method used - indication of the instrument is optional - and submit your results electronically to ESfEQA at [www.esfeqa.eu](http://www.esfeqa.eu).

Contact your local distributor of ESfEQA programs or ESfEQA directly if you need assistance for the registration, configuration and result submission in TEQA. Alternatively, though not preferred, use the fax

form provided on the ESfEQA website. In both cases indicate the reagent and method used for the analysis of the samples - indication of the instrument is optional.

In TEQA, all entries are confirmed by pressing the button "confirm". The successful transmission of the results is indicated by "data saved".

#### 6. Deadline for Data Submission

Deadline for data submission (time zone GMT +1):

Sample 1 and 2: 14/05/24

#### 7. Reports and Certificates

The data will be evaluated by ESfEQA.

The individual laboratory reports and certificates can be retrieved online at <https://teqa.esfeqa.eu>. The reports and certificates will be available within 3 weeks after the deadline of data submission.