ESfEQA
European Society for
External Quality Assessment

IFU_DENVAg, Ver003

Survey for Dengue Virus NS1 Antigen Testing (DENVAg)

EQA Provider:	ESfEQA GmbH Heidelberg
Survey Coordinator:	Dr. H. Depner

Instructions for Use

Notes:

The controls do not contain human material, but other biological materials. Therefore, the usual precautions in the laboratory for potentially hazardous material apply for these samples. The samples should be handled by trained personnel only.

Various tasks of the proficiency testing scheme may occasionally be given over to qualified subcontractors. However, ESfEQA is responsible to the participants for the subcontractor's work.

Results from sample analysis may only be disclosed to colleagues from other laboratories after the testing period has been concluded.

By registering for and participating in this EQA participants agree to accept the general terms and conditions of ESfEQA GmbH. These can be retrieved online at www.esfeqa.eu.

1. Intended Use

The samples are intended for use as qualitative control material for External Quality Assessment (EQA) in medical laboratories for the following parameter:

Dengue Virus NS1 Antigen

2. Product Description

The lyophilized samples simulate human serum specimens. Positive samples contain recombinant NS1 protein.

3. Storage and Stability

The samples should be stored upright at 2-8 °C. They are stable at least until the deadline for data submission as indicated on the sample labels.

4. Sample Preparation and Analysis

In order to ensure that all lyophilized sample material is at the bottom of the tube, briefly spin the tube with the lyophilized control (e.g. 30 seconds at 14.000 rpm) or tap the lyophilisate to the bottom of the tube before opening.

Reconstitute each sample by adding 0.3 mL of H_20 (preferably ddH_20) or diluent included in your test kit. Allow the control to reconstitute for 5 - 10 minutes, gently swirl and invert the closed tube to ensure homogeneity.

After reconstitution of the samples, they should be handled/treated like patient samples, and testing should be conducted in accordance with the instructions of the instrument and reagent manufacturers.

5. Dates and Submission of Test Results

Testing Periods: Please refer to the sample labels. Results can be submitted anytime within the testing period indicated on the sample labels.

For the evaluation of survey results, please indicate the reagent and method used - indication of the instrument is optional - and submit your results electronically to ESfEQA at https://teqa-labv2.esfeqa.eu.

Contact your local distributor of ESfEQA programs or ESfEQA directly if you need assistance for the registration, configuration and result submission in TEQA.

6. Deadline for Data Submission

The deadline for submitting results is indicated on the sample labels (time zone GMT +1).

7. Reports and Certificates

The data will be evaluated by ESfEQA. The individual laboratory reports and certificates can be retrieved online at https://teqa-labv2.esfeqa.eu.

ESfEQA GmbH
Siemensstr. 38
69123 Heidelberg
Germany

Phone: + 49 6221 4166-700
Fax: + 49 6221 4166-790
info@esfeqa.eu
www.esfeqa.eu